

**Product Information**

**Connection cable**



**Connection cable 1**  
for electrodes with S7-connector

Order no.	Length [m]	Protection class
S7-02	2	IP67
S7-05	5	IP67

**Connection cable 2**  
for electrodes with S7-connector at UNICON-pH, head mounting.

Order no.	Length [m]	Protection class
S7-K	-	-

**Connection cable 3**  
for electrodes with SMEK-VP-connector

Order no.	Length [m]	Protection class
KVP-02	2	IP67
KVP-05	5	IP67
KVP-10	10	IP67

(not used cables could be cutting)

**Connection cable 4**  
for electrodes with SMEK-VP-connector at UNICON-pH, head mounting.

Order no.	Length [m]	Protection class
KVP-K	-	-

**Calibrations tools**

**WTW technical buffer, 1000 ml with dosing container**

Order no.	PH-buffer value
TEP-4	4.01
TEP-7	7.00
TEP-10	10.00



**WTW ORP-buffer 250 ml bottle**

Order no.	Buffer value [mV]
RH28	427 (pH7)

**Storage solutions 250 ml DURAN-glass bottle, 3 mol KCL sterilized**

Order no.
pH-AL-250



**Cleaning solution 250 ml DURAN-glass bottle, Pepsin / hydrochloride acid**

Order no.
pH-RL-250

**Calibration container**  
for in-front-calibration material PMMA with level sign(20ml) and screw-cap PG13.5.  
(the pH-electrode must be screwed-in with the filled container).

Order no.
pH-KR-250



**Cleaning container 250 ml, to rinse the electrodes with water, material PP**

Order no.
pH-SB-250

